

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.541 °C	40.2 °C	-0.341 °C	± 0.5 °C	± 4.0E-01 °C

