

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 10.007 °C         | 10.3 °C   | 0.293 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.049 °C         | 24.8 °C   | -0.249 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 30.1 °C   | 0.038 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.541 °C         | 40.3 °C   | -0.241 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

