

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.272 °C	15.2 °C	-0.072 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.2 °C	-0.099 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.263 °C	30.3 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.243 %HR	69.6 %HR	-0.643 %HR	± 3.0 %HR	± 1.9E+00 %HR