

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49933	0.670	± 0.8	±	4.3E-01 mm						
0 a 1	0.99944	0.560	± 1	±	4.3E-01 mm						
0 a 1.5	1.49932	0.680	± 1.2	±	4.8E-01 mm						
0 a 2	1.99914	0.860	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49909	0.910	± 1.6	±	4.8E-01 mm						
0 a 3	2.99906	0.940	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49920	0.800	± 2	±	4.8E-01 mm						
0 a 4	3.99928	0.720	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49912	0.880	± 2.4	±	4.8E-01 mm						
0 a 5	4.99912	0.880	± 2.6	±	4.8E-01 mm						