

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
25 °C	25.249 °C	24.7 °C	-0.549 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.564 °C	29.3 °C	-1.264 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	38.0 %HR	2.317 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.0 %HR	-0.243 %HR	± 5.0 %HR	± 1.9E+00 %HR