

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-24.5 °C	0.539 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.121 °C	0.2 °C	0.321 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.9 °C	-0.093 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.9 °C	-0.107 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.998 °C	24.9 °C	-0.098 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.2 °C	-0.013 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.1 °C	-0.090 °C	± 0.5 °C	± 4.0E-01 °C

