

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.039 °C        | -25.7 °C  | -0.661 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.121 °C         | -0.5 °C   | -0.379 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 5.4 °C    | 0.407 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.007 °C         | 10.4 °C   | 0.393 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.998 °C         | 25.4 °C   | 0.402 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.213 °C         | 30.7 °C   | 0.487 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.190 °C         | 40.5 °C   | 0.310 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

