

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-25.2 °C	-0.161 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.121 °C	0.1 °C	0.221 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.4 °C	0.407 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.4 °C	0.393 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.998 °C	25.4 °C	0.402 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.213 °C	30.5 °C	0.287 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.6 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C

