

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.039 °C	-24.7 °C	0.339 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.251 °C	0.0 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.3 °C	0.307 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.3 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.6 °C	0.251 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.7 °C	0.136 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.3 °C	0.110 °C	± 0.5 °C	± 4.0E-01 °C

