

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	26.0 μm	2.70 μm	± 3 μm	±	5.8E-01 μm
51.6 μm	52.0 μm	0.40 μm	± 3 μm	±	5.8E-01 μm
124.6 μm	125.0 μm	0.40 μm	± 3 μm	±	5.9E-01 μm
250.3 μm	251.0 μm	0.70 μm	± 3 μm	±	6.3E-01 μm
486.00 μm	488.0 μm	2.00 μm	± 3 μm	±	7.5E-01 μm
963.00 μm	963.0 μm	0.00 μm	± 3 μm	±	1.1E+00 μm