

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.278 °C	10.8 °C	0.522 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.078 °C	32.5 °C	0.422 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	38.719 %HR	37.3 %HR	-1.419 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.071 %HR	73.1 %HR	-2.971 %HR	± 3.0 %HR	± 1.9E+00 %HR