

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.278 °C	11.3 °C	1.022 °C	± 1.0 °C	± 4.0E-01 °C
32 °C	32.078 °C	32.5 °C	0.422 °C	± 1.0 °C	± 4.0E-01 °C
40 %HR	38.719 %HR	41.8 %HR	3.081 %HR	± 3.5 %HR	± 1.9E+00 %HR
75 %HR	76.071 %HR	75.3 %HR	-0.771 %HR	± 3.5 %HR	± 1.9E+00 %HR