

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.278 °C	10.0 °C	-0.278 °C	± 1.0 °C	± 4.0E-01 °C
32 °C	32.078 °C	32.7 °C	0.622 °C	± 1.0 °C	± 4.0E-01 °C
40 %HR	38.719 %HR	42.1 %HR	3.381 %HR	± 3.5 %HR	± 1.9E+00 %HR
75 %HR	76.071 %HR	76.2 %HR	0.129 %HR	± 3.5 %HR	± 1.9E+00 %HR