

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.331 °C	-1.0 °C	-0.669 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.887 °C	9.0 °C	-0.887 °C	± 1.0 °C	± 4.0E-01 °C