

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.278 °C	10.0 °C	-0.278 °C	± 1.0 °C	± 4.0E-01 °C
32 °C	32.078 °C	32.0 °C	-0.078 °C	± 1.0 °C	± 4.0E-01 °C
40 %HR	38.719 %HR	42.0 %HR	3.281 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.071 %HR	73.0 %HR	-3.071 %HR	± 3.0 %HR	± 1.9E+00 %HR