

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.331 °C	-0.5 °C	-0.169 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	9.887 °C	9.5 °C	-0.387 °C	± 1.0 °C	± 4.0E-01 °C