

Certificado No. (Certificate #):		56268	Equipo (Instrument): Medidor de Espesores		
Fecha de Calibración (Calibration Date):		2025-04-25	Marca (Brand): Canned Instruments		
Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.0580 mm	0.0500 mm	-8.0000 µm	± µm	±	5.8E-01 µm
0.1060 mm	0.1008 mm	-5.2500 µm	± µm	±	5.8E-01 µm
0.2580 mm	0.2515 mm	-6.5000 µm	± µm	±	5.8E-01 µm
1.0050 mm	1.0048 mm	-0.2500 µm	± µm	±	5.8E-01 µm
2.0100 mm	2.0008 mm	-9.2500 µm	± µm	±	5.8E-01 µm
3.0040 mm	3.0123 mm	8.2500 µm	± µm	±	5.8E-01 µm
4.0070 mm	4.0145 mm	7.5000 µm	± µm	±	5.8E-01 µm
5.0030 mm	4.9935 mm	-9.5000 µm	± µm	±	5.8E-01 µm