

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Exterior / (Exterior)	0.0625 in	0.062 in	-0.001 in	± 0.0019 in	± 1.0E-03 in
Exterior / (Exterior)	0.500 in	0.501 in	0.001 in	± 0.0019 in	± 1.0E-03 in
Exterior / (Exterior)	1.000 in	1.001 in	0.001 in	± 0.0019 in	± 1.0E-03 in
Exterior / (Exterior)	3.000 in	3.001 in	0.001 in	± 0.0019 in	± 1.0E-03 in
Exterior / (Exterior)	6.000 in	6.001 in	0.001 in	± 0.0027 in	± 1.0E-03 in
Interior / (Interior)	0.500 in	0.500 in	0.000 in	± 0.0019 in	± 6.7E-04 in
Interior / (Interior)	5.000 in	5.000 in	0.000 in	± 0.0027 in	± 6.7E-04 in
Profundidad / (Depth)	0.500 in	0.500 in	0.000 in	± 0.0019 in	± 7.6E-04 in
Profundidad / (Depth)	5.000 in	5.000 in	0.000 in	± 0.0027 in	± 7.6E-04 in