

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.373 °C	30.4 °C	0.027 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.566 %HR	30.9 %HR	2.334 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	47.7 %HR	-1.858 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	73.4 %HR	-2.906 %HR	± 3.0 %HR	± 1.9E+00 %HR