

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.5 °C	-0.562 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.161 %HR	31.0 %HR	0.839 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	46.0 %HR	-3.998 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	65.0 %HR	-5.448 %HR	± 5.0 %HR	± 1.9E+00 %HR