

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
100 mm	100.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
101 mm	101.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
101.25 mm	101.25 mm	0 μm	± 6 μm	±	5.8E+00 μm
101.5 mm	101.50 mm	0 μm	± 6 μm	±	5.8E+00 μm
102 mm	102.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
103 mm	103.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
105 mm	105.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
110 mm	110.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
115 mm	115.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
120 mm	120.00 mm	0 μm	± 6 μm	±	5.8E+00 μm
125 mm	125.00 mm	0 μm	± 6 μm	±	5.8E+00 μm