

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
125 mm	125.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
126 mm	126.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
126.25 mm	126.25 mm	0 µm	± 6 µm	±	5.8E+00 µm
126.5 mm	126.50 mm	0 µm	± 6 µm	±	5.8E+00 µm
127 mm	127.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
128 mm	128.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
130 mm	130.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
135 mm	135.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
140 mm	140.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
145 mm	145.00 mm	0 µm	± 6 µm	±	5.8E+00 µm
150 mm	150.00 mm	0 µm	± 6 µm	±	5.8E+00 µm