

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.373 °C	30.3 °C	-0.073 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.566 %HR	31.1 %HR	2.534 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	51.1 %HR	1.542 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	73.6 %HR	-2.706 %HR	± 3.0 %HR	± 1.9E+00 %HR