

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.885 °C	1.5 °C	-0.385 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.753 °C	4.6 °C	-0.153 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.721 °C	7.4 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C