

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	2.375 in	2.38 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.400 in	2.40 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.430 in	2.43 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.460 in	2.46 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.500 in	2.50 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.530 in	2.53 in	0 in	± 0.0019685 in	± 5.8E-01 in