

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	2.375 in	2.38 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.400 in	2.40 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.430 in	2.43 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.460 in	2.46 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.500 in	2.50 in	0 in	± 0.0019685 in	± 5.8E-01 in
Profundidad / (Depth)	2.530 in	2.53 in	0 in	± 0.0019685 in	± 5.8E-01 in