

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1 mm              | 0.96 mm   | -40 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 2 mm              | 1.99 mm   | -10 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 5 mm              | 5.05 mm   | 50 µm         | ± 10 µm           | ±                           | 5.8E+00 µm |
| 10 mm             | 10.10 mm  | 100 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 15 mm             | 15.20 mm  | 200 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 20 mm             | 20.39 mm  | 390 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 25 mm             | 25.58 mm  | 580 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 50 mm             | 50.63 mm  | 630 µm        | ± 10 µm           | ±                           | 5.8E+00 µm |
| 100 mm            | 98.85 mm  | -1150 µm      | ± 10 µm           | ±                           | 5.8E+00 µm |