

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	45 mm	45.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	90 mm	90.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	135 mm	135.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	180 mm	180.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	230 mm	230.03 mm	30 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	255 mm	255.02 mm	20 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	300 mm	300.03 mm	30 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	355 mm	355.02 mm	20 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	400 mm	400.03 mm	30 µm	± 40 µm	± 6.1E+00 µm