

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.373 °C	30.7 °C	0.327 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.566 %HR	31.5 %HR	2.934 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	49.5 %HR	-0.058 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.306 %HR	73.0 %HR	-3.306 %HR	± 3.0 %HR	± 1.9E+00 %HR