

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	30.4 °C	-0.063 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	34.6 %HR	-0.469 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	51.4 %HR	1.218 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	68.1 %HR	-1.887 %HR	± 3.0 %HR	± 1.9E+00 %HR