

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.00	mm	1.001	mm	1 μm	± 6 μm	±	5.9E-01 μm
1.25	mm	1.250	mm	0 μm	± 4 μm	±	5.8E-01 μm
1.50	mm	1.501	mm	1 μm	± 4 μm	±	5.8E-01 μm
2.00	mm	2.001	mm	1 μm	± 4 μm	±	5.8E-01 μm
3.00	mm	3.001	mm	1 μm	± 4 μm	±	5.8E-01 μm
5.00	mm	5.002	mm	2 μm	± 4 μm	±	5.8E-01 μm
8.00	mm	8.002	mm	2 μm	± 4 μm	±	5.9E-01 μm
10.00	mm	10.001	mm	1 μm	± 4 μm	±	5.9E-01 μm
15.00	mm	15.001	mm	1 μm	± 4 μm	±	5.8E-01 μm
20.00	mm	20.001	mm	1 μm	± 4 μm	±	5.9E-01 μm
25.00	mm	25.002	mm	2 μm	± 4 μm	±	5.9E-01 μm