

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.463 °C	29.8 °C	-0.663 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	36.0 %HR	0.931 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	47.0 %HR	-3.182 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	63.0 %HR	-6.987 %HR	± 5.0 %HR	± 1.9E+00 %HR