

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	29.7 °C	-0.763 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	36.0 %HR	0.931 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	46.0 %HR	-4.182 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.987 %HR	61.0 %HR	-8.987 %HR	± 5.0 %HR	± 1.9E+00 %HR