

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.221 °C	0.1 °C	0.321 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.723 °C	16.0 °C	0.277 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	35.8 %HR	0.731 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	49.1 %HR	-0.898 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	70.7 %HR	0.355 %HR	± 3.0 %HR	± 1.9E+00 %HR