

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.221 °C	-0.5 °C	-0.279 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.723 °C	15.8 °C	0.077 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.5 °C	0.438 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	35.1 %HR	0.031 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	50.3 %HR	0.302 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	71.2 %HR	0.855 %HR	± 3.0 %HR	± 1.9E+00 %HR