

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.221 °C	-0.5 °C	-0.279 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.723 °C	15.4 °C	-0.323 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.5 °C	0.438 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	35.1 %HR	0.031 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	49.7 %HR	-0.298 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	71.1 %HR	0.755 %HR	± 3.0 %HR	± 1.9E+00 %HR