

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.221 °C	-0.2 °C	0.021 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.723 °C	16.0 °C	0.277 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.069 %HR	36.3 %HR	1.231 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	50.7 %HR	0.702 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	70.3 %HR	-0.045 %HR	± 3.0 %HR	± 1.9E+00 %HR