

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99943	0.570	± 1	± 6.6E-01 mm					
0 a 2	2.00003	-0.030	± 1.4	± 7.0E-01 mm					
0 a 3	2.99926	0.740	± 1.8	± 7.0E-01 mm					
0 a 4	3.99958	0.420	± 2.2	± 7.0E-01 mm					
0 a 5	4.99955	0.450	± 2.6	± 7.0E-01 mm					
0 a 6	5.99978	0.220	± 3	± 6.1E-01 mm					
0 a 7	6.99943	0.570	± 3.4	± 6.1E-01 mm					
0 a 8	7.99973	0.270	± 3.8	± 6.1E-01 mm					
0 a 9	8.99941	0.590	± 4.2	± 6.1E-01 mm					
0 a 10	9.99952	0.480	± 4.6	± 6.1E-01 mm					
0 a 11	10.99972	0.280	± 5	± 6.1E-01 mm					
0 a 12	12.00005	-0.050	± 5.4	± 6.1E-01 mm					
0 a 13	12.99983	0.170	± 5.8	± 6.1E-01 mm					
0 a 14	14.00022	-0.220	± 6.2	± 6.1E-01 mm					
0 a 15	14.99992	0.080	± 6.6	± 6.1E-01 mm					
0 a 16	16.00015	-0.150	± 7	± 6.1E-01 mm					
0 a 17	17.00002	-0.020	± 7.4	± 6.1E-01 mm					
0 a 18	18.00045	-0.450	± 7.8	± 6.1E-01 mm					
0 a 19	19.00018	-0.180	± 8.2	± 6.1E-01 mm					
0 a 20	20.00057	-0.570	± 8.6	± 6.1E-01 mm					
0 a 21	21.00020	-0.200	± 9	± 7.4E-01 mm					
0 a 22	22.00068	-0.680	± 9.4	± 7.4E-01 mm					
0 a 23	23.00023	-0.230	± 9.8	± 7.4E-01 mm					
0 a 24	24.00050	-0.500	± 10.2	± 7.4E-01 mm					
0 a 25	25.00032	-0.320	± 10.6	± 7.4E-01 mm					
0 a 26	26.00039	-0.390	± 11	± 7.4E-01 mm					
0 a 27	27.00043	-0.430	± 11.4	± 7.4E-01 mm					
0 a 28	28.00066	-0.660	± 11.8	± 7.4E-01 mm					
0 a 29	29.00035	-0.350	± 12.2	± 7.4E-01 mm					
0 a 30	30.00090	-0.900	± 12.6	± 7.4E-01 mm					