

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49906	0.940	± 1.6	± 4.3E-01 mm					
0 a 1	0.99914	0.860	± 2	± 4.3E-01 mm					
0 a 1.5	1.49944	0.560	± 2.4	± 4.8E-01 mm					
0 a 2	1.99917	0.830	± 2.8	± 4.8E-01 mm					
0 a 2.5	2.49915	0.850	± 3.2	± 4.8E-01 mm					
0 a 3	2.99909	0.910	± 3.6	± 4.8E-01 mm					
0 a 3.5	3.49958	0.420	± 4	± 4.8E-01 mm					
0 a 4	3.99972	0.280	± 4.4	± 4.8E-01 mm					
0 a 4.5	4.49923	0.770	± 4.8	± 4.8E-01 mm					
0 a 5	4.99974	0.260	± 5.2	± 4.8E-01 mm					