

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	0.99 mm	-10.00 µm	± 0.015 µm	±	5.8E+00 µm
2.00 mm	1.99 mm	-10.00 µm	± 0.020 µm	±	5.8E+00 µm
5.00 mm	4.99 mm	-10.00 µm	± 0.035 µm	±	5.8E+00 µm
10.00 mm	9.99 mm	-10.00 µm	± 0.060 µm	±	5.8E+00 µm
15.00 mm	14.99 mm	-10.00 µm	± 0.085 µm	±	5.8E+00 µm
20.00 mm	19.99 mm	-10.00 µm	± 0.11 µm	±	5.8E+00 µm
25.00 mm	24.99 mm	-10.00 µm	± 0.14 µm	±	5.8E+00 µm
50.00 mm	49.99 mm	-10.00 µm	± 0.26 µm	±	5.8E+00 µm
100.00 mm	100.00 mm	0.00 µm	± 0.51 µm	±	5.8E+00 µm