

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.030 °C	0.4 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.356 °C	20.7 °C	0.344 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.146 °C	34.9 °C	-0.246 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.801 %HR	31.7 %HR	2.899 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	54.261 %HR	52.8 %HR	-1.461 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.151 %HR	77.3 %HR	-2.851 %HR	± 3.0 %HR	± 1.9E+00 %HR