

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 0 °C                       | -0.030 °C                  | 0.0 °C             | 0.030 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.356 °C                  | 20.3 °C            | -0.056 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 35 °C                      | 35.146 °C                  | 34.9 °C            | -0.246 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 28.801 %HR                 | 31.7 %HR           | 2.899 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 55 %HR                     | 54.261 %HR                 | 55.7 %HR           | 1.439 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 80 %HR                     | 80.151 %HR                 | 79.4 %HR           | -0.751 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |