

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.423 °C	30.0 °C	-0.423 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.156 °C	34.8 °C	-0.356 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	49.149 %HR	45.0 %HR	-4.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.946 %HR	61.0 %HR	-8.946 %HR	± 5.0 %HR	± 1.9E+00 %HR