

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.423 °C	29.8 °C	-0.623 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	34.0 %HR	4.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.149 %HR	47.0 %HR	-2.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.946 %HR	62.0 %HR	-7.946 %HR	± 5.0 %HR	± 1.9E+00 %HR