

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.423 °C	29.9 °C	-0.523 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	32.0 %HR	2.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.149 %HR	49.0 %HR	-0.149 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.946 %HR	68.0 %HR	-1.946 %HR	± 5.0 %HR	± 1.9E+00 %HR