

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
25 mm	25.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
25.25 mm	25.25 mm	0 μm	± 13 μm	±	5.8E+00 μm
25.5 mm	25.50 mm	0 μm	± 13 μm	±	5.8E+00 μm
26 mm	26.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
26 mm	26.25 mm	0 μm	± 13 μm	±	5.8E+00 μm
27 mm	26.50 mm	0 μm	± 13 μm	±	5.8E+00 μm
27 mm	27.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
28 mm	28.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
29 mm	29.00 mm	0 μm	± 13 μm	±	5.8E+00 μm