

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1 mm	1.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
1.25 mm	1.25 mm	0 μm	± 13 μm	±	5.8E+00 μm
1.5 mm	1.50 mm	0 μm	± 13 μm	±	5.8E+00 μm
2 mm	2.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
3 mm	3.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
5 mm	5.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
6 mm	6.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
8 mm	8.00 mm	0 μm	± 13 μm	±	5.8E+00 μm
10 mm	10.00 mm	0 μm	± 13 μm	±	5.8E+00 μm