

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.248 °C	10.5 °C	0.252 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	20.2 °C	0.175 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.320 °C	29.0 °C	-0.320 °C	± 0.5 °C	± 4.0E-01 °C
67 %HR	68.188 %HR	69.17 %HR	0.982 %HR	± 5.0 %HR	± 1.9E+00 %HR
88 %HR	90.356 %HR	85.98 %HR	-4.376 %HR	± 5.0 %HR	± 1.9E+00 %HR