

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.579 °C	10.7 °C	0.121 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.727 °C	20.7 °C	-0.027 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.200 °C	29.0 °C	-0.200 °C	± 0.5 °C	± 4.0E-01 °C
67 %HR	67.002 %HR	66.12 %HR	-0.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
88 %HR	88.311 %HR	90.27 %HR	1.959 %HR	± 5.0 %HR	± 1.9E+00 %HR