

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.599 °C	11.0 °C	0.401 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.707 °C	20.9 °C	0.193 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.210 °C	29.0 °C	-0.210 °C	± 0.5 °C	± 4.0E-01 °C
67 %HR	67.206 %HR	66.46 %HR	-0.746 %HR	± 5.0 %HR	± 1.9E+00 %HR
88 %HR	88.433 %HR	93.43 %HR	4.997 %HR	± 5.0 %HR	± 1.9E+00 %HR